

2014 Rapid DNA Maturity Assessment Results

Erica Romsos

National Institute of Standards and Technology

Global Identity Summit

Tampa, FL September 22, 2015





Disclaimer

We will mention commercial STR kit and instrument names, but we are in no way attempting to endorse any specific products.

NIST Disclaimer: Certain commercial equipment, instruments and materials are identified in order to specify experimental procedures as completely as possible. In no case does such identification imply a recommendation or it imply that any of the materials, instruments or equipment identified are necessarily the best available for the purpose.


Information presented does not necessarily represent the official position of the National Institute of Standards and Technology or the U.S. Department of Justice.

Purpose of Maturity Assessment

- To assess the status in the fall 2014 of rapid DNA typing technology for the CODIS 13 core loci
 - In support of lab use and future external (nonlab-based) Rapid DNA implementation
- Integrated (swab in allele detection)
 instruments capable of genotyping the
 core CODIS 13 STR markers were eligible
 for the study

Rapid DNA Instruments

ANDE (NetBio)



- One biochipset
 - Stored at RT
 - Shelf life ≈ 6 months
- RFID swabs tagged for sample tracking

PowerPlex 16 loci ≈86 min runtime (5 samples)

ANDE PP16

RapidHIT 200 (IntegenX)




- Kit = 4 components
 - Stored between RT-4°C
 - Shelf life ≈ 6 months @ 4°C
- Cotton Swabs

PowerPlex 16 loci ≈90 min runtime (5 samples) GlobalFiler Express loci ≈120 min runtime (1-7 samples)

RH200 PP16

RH200 GFE

R-DNA Maturity Assessment



Timeline of Maturity Assessment

January 2014: Buccal samples collected at NIST and stored at RT

October 2014: Samples shipped to participating laboratories

October-December 2014: Data generated and electronically returned to NIST

November-December 2014: Data analyzed at NIST

Maturity Assessment

Participating Independent Total Samples Instrument Chemistry Instruments (11) Tested (280) Laboratories (7) Platforms (2) NetBio ANDE **Federal** 5 PowerPlex 16 100 State IntegenX RapidHIT 200 PowerPlex 16 60 6 GlobalFiler 120 Private **Express**

NIST Analysis Parameters

Rapid DNA Analysis: Without human intervention


 Modified Rapid DNA Analysis: Expert interpretation and analysis of electropherogram

https://www.fbi.gov/about-us/lab/biometric-analysis/codis/rapid-dna-addendum-to-qas-final-effective-12-1-2014

- Additional analysis (PHR, Stutter, etc.) of the data performed with GeneMapper IDX v 1.3
 - Custom bins and panels designed for analysis of all data in GeneMapper
- In-house Excel programs used to analyze peak height ratios, stutter, and precision


Overall Success

Success was measured by **complete and concordant genotypes** produced by the integrated rapid DNA devices as compared to lab generated correct genotypes




Locus Success (Rapid Analysis)

PowerPlex 16



GlobalFiler Express



RH200 GFE: n=88

Peak Height Ratios


PowerPlex 16

Locus	Median		
Penta_E	0.81		
AMEL	0.83		
Penta_D	0.84		
D18S51	0.86		
D3S1358	0.87		
D8S1179	0.87		
TPOX	0.87		
D5S818	0.88		
∨WA	0.88		
D21S11	0.88		
D16S539	0.88		
D13S317	0.89		
CSF1PO	0.89		
FGA	0.89		
D7S820	0.89		
TH01	0.93		

Full Profiles: n=118

Peak height ratios were calculated for all **complete profiles** for the PowerPlex 16 and GlobalFiler Express chemistries.

The PowerPlex 16 data is a <u>combination</u> of the data generated from both ANDE and the RapidHIT 200.



GlobalFiler Express

EXPICOS					
Locus	Median				
SE33	0.79				
D2S1338	0.82				
D5S818	0.85				
D18S51	0.85				
D12S391	0.86				
D21S11	0.87				
CSF1PO	0.87				
∨WA	0.88				
D7S820	0.88				
TPOX	0.89				
D16S539	0.89				
D1S1656	0.89				
D22S1045	0.89				
D8S1179	0.90				
D13S317	0.90				
AMEL	0.90				
D3S1358	0.90				
D19S433	0.90				
D10S1248	0.91				
TH01	0.91				
FGA	0.92				
D2S441	0.92				

Full Profiles: n=67

Stutter Percentage


PowerPlex 16

•	B. 0 12
Locus	Median
Penta_D	1.47
TH01	2.28
TPOX	2.82
Penta_E	4.12
D7S820	5.76
D13S317	6.48
D18S51	7.15
D8S1179	7.30
CSF1PO	7.52
D16S539	7.67
D5S818	8.36
FGA	8.78
∨WA	9.30
D3S1358	10.23
D21S11	10.72

Stutter percentages were calculated for all **complete profiles** for the PowerPlex 16 and GlobalFiler Express chemistries.

The PowerPlex 16 data is a **combination of the data generated** from both ANDE and the RapidHIT 200.

All stutter percentages within observed kit manufacturer developmental validation range



Full Profiles: n=118

GlobalFiler Express

Loci	Median		
TH01	1.27		
TPOX	3.64		
D7S820	4.75		
D2S441	4.76		
DYS391	5.73		
D16S539	5.84		
D13S317	5.90		
CSF1PO	6.12		
D8S1179	6.59		
D18S51	6.67		
D5S818	6.76		
D22S1045	7.00		
D19S433	7.10		
FGA	7.20		
D3S1358	8.31		
D10S1248	8.36		
D21S11	8.60		
D2S1338	8.71		
D1S1656	8.77		
√WA	9.28		
D12S391	9.46		
SE33	15.56		


Full Profiles: n=67

Ladders


- Both instruments run a ladder with each run
 - Incorporated into the chip/kit
- Each instrument contains an "onboard" ladder(s)
 - For use if the ladder on the chip fails
- Ladders and internal size standard allow for accurate allele calling
 - Poor precision (>0.5 bp) can result in miscalled data

Ladder Precision

PowerPlex 16



GlobalFiler Express



Allelic Precision of 0.125 bp

Ladders: : n=46

Allelic Precision of 0.133 bp

Ladders: : n=74

Maturity Assessment Summary

- 11 instruments within 7 laboratories tested
 - Total of 280 samples examined
- Data generated October-December 2014 and returned to NIST
- Changes since 2014 Maturity Assessment Data was generated (between both companies)
 - Known changes to manufacturing, software, and hardware

Summary of Results

- 2014 R-DNA Maturity Assessment exhibited a 76% success rate for the CODIS 13 Core Loci using Rapid DNA Analysis
 - Success ranged from 45% to 85% across laboratories, chemistries, and instruments
- Precision is below 0.25 bp on for both PP16 and GFE data generated
- Continuing to run R-DNA platforms with newer chemistries and upgrades

Final Results

Rapid DNA Instrument Platforms	Number of Participating Labs	Total Instruments	Samples Attempted	Core CODIS Success (Rapid DNA Analysis)	Core CODIS Success (Modified Rapid DNA Analysis)
2	7	11	280	76.1%	80.0%

Overall success for the R-DNA maturity assessment is reported:

http://www.nist.gov/mml/bmd/genetics/dna_biometrics.cfm



Rapid DNA Maturity Assessment



NUST

habert bribe ofference are betrology
Detrology foretherate, U.S. Department Commence

Email:
Erica.Romsos@nist.gov

Erica L. Romsos¹, Sanae Lembirick², and Peter M. Vallone¹

1 U.S. National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899-8314, USA 2 Montgomery College, Rockville, MD 20850, USA



Poster presented at the 26th Congress of the International Society for Forensic Genetics (Krakow, Poland), September 2-5, 2015

http://www.cstl.nist.gov/strbase/pub_pres/RomsosISFG2015RapidDNA.pdf



Thanks to David Duewer and Sanae Lembirik for assistance with data analysis

Contact Info:

erica.romsos@nist.gov 301-975-5107

Funding

Performance Evaluation

DHS – Rapid DNA **FBI** - the Evaluation of Prototype and Kinship Forensic DNA Typing as a Biometric Tool



