

2014 Rapid DNA Maturity Assessment Results

Erica Romsos

National Institute of Standards and Technology

Potomac Regional Symposium on Forensic DNA Analysis

Richmond, VA November 2, 2015



Disclaimer

We will mention commercial STR kit and instrument names, but we are in no way attempting to endorse any specific products.

NIST Disclaimer: Certain commercial equipment, instruments and materials are identified in order to specify experimental procedures as completely as possible. In no case does such identification imply a recommendation or it imply that any of the materials, instruments or equipment identified are necessarily the best available for the purpose.

Information presented does not necessarily represent the official position of the National Institute of Standards and Technology or the U.S. Department of Justice.

Previous Maturity Assessments

- Conducted Summer 2013
- Presented at the 2013 Global Identity Summit
- Two R-DNA developers
- Three testing sites
- A total of 350 reference buccal swabs run
- Success defined as the automated calling of the 13 core STR loci
- Overall success = 87.4%

2014 Maturity Assessment

- Purpose to assess the status in the fall 2014 of rapid DNA typing technology for the CODIS 13 core loci
 - In support of lab use and future external (nonlab-based) Rapid DNA implementation
- Integrated (swab in allele detection) instruments capable of genotyping the core CODIS 13 STR markers were eligible for the study

Rapid DNA Instruments ANDE (NetBio) RapidHIT 200 (IntegenX)



- One biochipset
 - Stored at RT
 - Shelf life \approx 6 months
- RFID swabs tagged for sample tracking

PowerPlex 16 loci ≈86 min runtime (5 samples)

ANDE PP16



- Kit = 4 components •
 - Stored between RT-4°C
 - Shelf life ≈ 6 months @ 4°C
- Cotton Swabs

PowerPlex 16 loci ≈90 min runtime (5 samples)

GlobalFiler Express loci ≈120 min runtime (1-7 samples)

RH200 PP16

RH200 GFE

ANDE (NetBio)



ANDE (NetBio)



ANDE-Successful Profile



RapidHIT 200 (IntegenX)



RapidHIT 200 (IntegenX)

Kit = 4 components Stored at 4° C Shelf life \approx 3 months Cotton swabs

Running PP16 loci ≈108 min runtime

Running GFE loci ≈180 min runtime



RapidHIT 200-Correct Profile



2014 R-DNA Maturity Assessment



Timeline of Maturity Assessment

January 2014: Buccal samples collected at NIST and stored at RT

October 2014: Samples shipped to participating laboratories

October-December 2014: Data generated and electronically returned to NIST

November-December 2014: Data analyzed at NIST

20 Swabs provided

Maturity Assessment



Analysis: FBI Definitions

- Rapid DNA Analysis: describes the fully automated (hands free) process of developing a CODIS Core STR profile from a reference sample buccal swab. The "swab in – profile out" process consists of automated extraction, amplification, separation, detection and allele calling without human intervention.
- Modified Rapid DNA Analysis: describes the automated process of developing a CODIS Core STR profile from a reference sample buccal swab. This process consists of integrated extraction, amplification, separation, detection without human intervention, but requires human interpretation and technical review.

D3S1358	TH01	D21S11	D18S51	PentaE	D5S818	D13S317	D7S820	D16S539	CSF1PO	PentaD	AMEL	vWA	D8S1179	трох	FGA



D3S1358	TH01	D21S11	D18S51	PentaE	D5S818	D13S317	D7S820	D16S539	CSF1PO	PentaD	AMEL	vWA	D8S1179	трох	FGA

D3S1358	TH01	D21S11	D18S51	PentaE	D5S818	D13S317	D7S820	D16S539	CSF1PO	PentaD	AMEL	vWA	D8S1179	трох	FGA

Rapid DNA Analysis: Failure



D3S1358	TH01	D21S11	D18S51	PentaE	D5S818	D13S317	D7S820	D16S539	CSF1PO	PentaD	AMEL	vWA	D8S1179	ΤΡΟΧ	FGA

NIST Analysis Parameters

- Additional analysis (PHR, Stutter, etc.) of the data performed with GeneMapper IDX v 1.3
 - Custom bins and panels designed for analysis of all data in GeneMapper IDX v1.3 for both ANDE and RapidHIT 200
- In-house Excel programs used to analyze peak height ratios, stutter, and precision

Overall Success

Success was measured by **complete and concordant genotypes** produced by the integrated rapid DNA devices as compared to lab generated correct genotypes





Peak Height Ratios

PowerPlex 16

Locus	Median
Penta_E	0.81
AMEL	0.83
Penta_D	0.84
D18S51	0.86
D3S1358	0.87
D8S1179	0.87
TPOX	0.87
D5S818	0.88
√WA	0.88
D21S11	0.88
D16S539	0.88
D13S317	0.89
CSF1PO	0.89
FGA	0.89
D7S820	0.89
TH01	0.93

Peak height ratios were calculated for all **complete profiles** for the PowerPlex 16 and GlobalFiler Express chemistries.

The PowerPlex 16 data is a <u>combination</u> <u>of the data generated</u> from both ANDE and the RapidHIT 200.



Express Median Locus 0.79 **SE33** D2S1338 0.82 D5S818 0.85 D18S51 0.85 D12S391 0.86 D21S11 0.87 CSF1PO 0.87 0.88 **WWA** D7S820 0.88 TPOX 0.89 D16S539 0.89 D1S1656 0.89 D22S1045 0.89 D8S1179 0.90 D13S317 0.90 AMEL 0.90 D3S1358 0.90 D19S433 0.90 D10S1248 0.91 TH01 0.91

GlobalFiler

Full Profiles: n=118

Full Profiles: n=67

FGA

D2S441

0.92

0.92

Stutter Percentage

Stutter percentages were calculated for all **complete profiles** for the PowerPlex 16 and GlobalFiler Express chemistries.

The PowerPlex 16 data is a <u>combination</u> of the data generated from both ANDE and the RapidHIT 200.



GlobalFiler Express

Loci	Median
TH01	1.27
TPOX	3.64
D7S820	4.75
D2S441	4.76
DYS391	5.73
D16S539	5.84
D13S317	5.90
CSF1PO	6.12
D8S1179	6.59
D18S51	6.67
D5S818	6.76
D22S1045	7.00
D19S433	7.10
FGA	7.20
D3S1358	8.31
D10S1248	8.36
D21S11	8.60
D2S1338	8.71
D1S1656	8.77
VWA	9.28
D12S391	9.46
SE33	15.56

Full Profiles: n=67

PowerPlex 16 Locus Median 1.47 Penta D TH01 2.28 2.82 TPOX 4.12 Penta E 5.76 D7S820 D13S317 6.48 7.15 D18S51 D8S1179 7.30 CSF1PO 7.52 D16S539 7.67 8.36 D5S818 FGA 8.78 9.30 VWA D3S1358 10.23 D21S11 10.72

Full Profiles: n=118

Ladders

- Both instruments run a ladder with each run
 - Incorporated into the chip/kit
- Each instrument contains an "onboard" ladder(s)
 - For use if the ladder on the chip fails
- Ladders and internal size standard allow for accurate allele calling
 - Poor precision (>0.5 bp) can result in miscalled data



IntegenX Ladders



Ladder Precision



Allelic Precision of 0.125 bp

Ladders: : n=46

Allelic Precision of 0.133 bp

Ladders: : n=74

Maturity Assessment Summary

- 11 instruments within 7 laboratories tested
 Total of 280 samples examined
- Data generated October-December 2014 and returned to NIST
- Changes since 2014 Maturity Assessment Data was generated (between both companies)
 - Known changes to manufacturing, software, and hardware

Summary of Results

- 2014 R-DNA Maturity Assessment exhibited a 76% success rate for the CODIS 13 Core Loci using Rapid DNA Analysis
 - Success ranged from 45% to 85% across
 laboratories, chemistries, and instruments
- Precision is below 0.25 bp on for both PP16 and GFE data generated
- Continuing to run R-DNA platforms with newer chemistries and upgrades

Final Results Available Online

Rapid DNA Instrument Platforms	Number of Participating Labs	Total Instruments	Samples Attempted	Core CODIS Success (Rapid DNA Analysis)	Core CODIS Success (Modified Rapid DNA Analysis)
2	7	11	280	76.1%	80.0%

Overall success for the R-DNA maturity assessment is reported:

http://www.nist.gov/mml/bmd/genetics/dna_biometrics.cfm



haderal bedeen officerative are betredays betredays templatistics, tak Departmented Connector Email:

Erica.Romsos@nist.gov

Rapid DNA Maturity Assessment

Erica L. Romsos¹, Sanae Lembirick², and Peter M. Vallone¹ ¹U.S. National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899-8314, USA ² Montgomery College, Rockville, MD 20850, USA



Poster presented at the 26th Congress of the International Society for Forensic Genetics (Krakow, Poland), September 2-5, 2015 http://www.cstl.nist.gov/strbase/pub_pres/RomsosISFG2015RapidDNA.pdf

Thank you for your attention!

Thanks to David Duewer and Sanae Lembirik for assistance with data analysis

Contact Info:

erica.romsos@nist.gov

301-975-5107

Funding

Performance Evaluation

DHS – Rapid DNA **FBI** - the Evaluation of Prototype and Kinship Forensic DNA Typing as a Biometric Tool



