


Application of Emerging Technologies at NIST

GenomeID Forum – Emerging Forensic Genomic Applications
Center of Advanced Forensic DNA Analysis
September 9, 2014 - Greenville, NC

Peter M. Vallone, Ph.D.
Applied Genetics Group
Biomolecular Measurement Division






Disclaimer


NIST Disclaimer: Certain commercial equipment, instruments and materials are identified in order to specify experimental procedures as completely as possible. In no case does such identification imply a recommendation or it imply that any of the materials, instruments or equipment identified are necessarily the best available for the purpose.

Information presented does not necessarily represent the official position of the National Institute of Standards and Technology or the U.S. Department of Justice.

Our group receives or has received funding from the FBI Laboratory and the National Institute of Justice.

Outline

- NIST forensic SRMs 
- Digital PCR 
- Next-generation sequencing 



Current Characterization of Forensic SRMs

- **2391c PCR Based DNA profiling standard**
 - 68 STR markers (51 autosomal + 17 Y chromosome)
 - STR repeat lengths (alleles) were certified using multiple (unique) PCR primer sets
 - Sanger sequencing was only performed for loci without multiple PCR primer sets (**only 10%**)
- **2392 & 2392-I Mitochondrial DNA sequencing standard**
 - Entire mtGenome (≈16,569 bp) was certified by Sanger sequencing
- **2372 Human DNA Quantitation Standard**
 - UV absorbance (decadic attenuation) measurement




Goal: Characterize Existing Forensic SRMs with New and Emerging Technologies

- **SRM 2391c: Certify sequence information for STR loci**
 - Sanger and NGS methods
 - Supports adoption of NGS in forensic community
 - Understand bias inherent to specific NGS platforms: chemistry and bioinformatics
- **SRMs 2392 and 2392-I: confirm Sanger data with high coverage NGS methods**
 - Detect lower level heteroplasmies (<20 %)
- **SRM 2372: certify concentration with an absolute PCR-based method**
 - Digital PCR provides this capability


Certified, Reference & Information Values

- Certified Value**
 - NIST has highest confidence in accuracy
 - All known/suspected sources of bias investigated/taken into account
Two or more methods e.g. Sanger sequencing AND genotyping with multiple primer sets
- Reference Value**
 - Best estimate of true value
 - All possible sources of bias NOT fully investigated by NIST
Genotyping with only two sets of primers
- Information Value**
 - Of interest and use to SRM user
 - Insufficient information available to assess uncertainty of value
Genotyping with only one set of primers

Outline

- NIST forensic SRMs 
- Digital PCR 
- Next-generation sequencing 

SRM 2372 DNA Quantitation Standard

- Used for calibrating DNA quantitation standards
 - (qPCR ktis)
- Current stock: 31 month supply 
- In the process of preparing SRM 2372a
- Characterize with dPCR versus UV absorbance

Digital PCR (dPCR) Overview

- A sample is partitioned so that individual nucleic acid targets within the sample are localized
 - Microfluidic (Fluidigm BioMark)
 - Emulsion/droplet PCR (Bio-Rad QX100, RainDance)
- Each partition will contain a negative or positive PCR reaction
- Nucleic acid targets may be quantified by counting the regions that contain PCR end-product
 - A standard curve is not required



• Sykes, P.J. et al. AA (1992) "Quantitation of targets for PCR by use of limiting dilution". Biotechniques 13 (3): 444-449
 • Kalinina, O et al. (1997) "Nanoliter scale PCR with TaqMan detection". Nucleic Acids Research 25 (10): 1999-2004
 • Vogelsberg and Klotzer (1998) "Digital PCR". Proc Natl Acad Sci U S A, 95 (18): 9236-9241
 • Pohl and Shih (2004) "Principle and applications of digital PCR". Expert Rev Mol Diagn 4 (1): 41-47
 • Dressman et al. (2003). "Transforming single DNA molecules into fluorescent magnetic particles for detection and enumeration of genetic variations". Proc Natl Acad Sci USA 100 (15): 8817-8822

Fluidigm BioMark

12.765


- Fluidic module transfers PCR mastermix onto chip
- 'Reader' performs thermal cycling and fluorescence detection (real-time PCR)

Fluidigm Digital Arrays
 12.765 = 765 chambers x 12 panels (samples)
 48.770 = 770 chambers x 48 panels (samples)

- Well volumes
 - 6 nL (12 sample)
 - 0.85 nL (48 samples)
- TaqMan compatible chemistry
- FAM-VIC dye detection

Fluorescent signal as a function of amplification cycle in 765 dPCR reactions









Majority of the wells amplify within a narrow range of C_T values

Later amplification may be due to:
 Damaged target
 Partially blocked target
 Secondary binding sites

Grey lines are no amplification

$$\text{Concentration (copies per microliter)} = \frac{\text{total number of wells} \cdot \ln \left(\frac{\text{total number of wells}}{\text{total number of negative wells}} \right)}{\text{volume of all PCR reactions (microliters)}}$$


Number of wells with signal relates to the number of copies of starting DNA 






Design and validate multiple dPCR assays for certification of SRM 2372a

Convert copies/µL and calculate the DNA concentration as ng/µL:

Assay	Chromosome	Average ng/µL	sd
D6S474	6	56.3	1.2
D9S2157	9	55.2	0.7
HBB1	11	51.7	1.5
D5S2500	5	51.7	0.9
D14S1434	14	52.1	0.4
2PR4	2	50.6	0.9
22C3	22	50.0	1.2
EIF5	2	49.0	0.3
D1P32.3	1	39.7	0.8
Average of all (except D1P32.3)		52.1	2.5




Outline


- NIST forensic SRMs
 
- Digital PCR
 
- Next-generation sequencing
 

Platforms

- Illumina**
 - MiSeq/FGx
 - HiSeq 2000/2500
 - GAIx
 - NextSeq 500
 - HiSeq X Ten
- Life Technologies**
 - SOLiD (5500 series)
 - Ion Torrent PGM
 - Ion Torrent Proton
- Pacific Biosciences**
 - PACBIO RS II
- 454 Roche**
 - GS jr
 - GS FLX+
- Oxford Nanopore**
Qiagen
Others on the horizon


Characterization of SRM 2392 and 2392-1 Mitochondrial genome sequencing standard





- Sequence the entire mitochondrial genome
 - Two, three and twelve amplicon strategies
- PGM, MiSeq, HiSeq, and SOLiD platforms
- Check concordance with Sanger results
- Detection of heteroplasmy (< 20%)
- Issues with homopolymers
 - Chemistry and informatics related

NGS Support for mtDNA Analysis

- PGM & MiSeq analysis
 - 5% SNP calling threshold
 - Concordance across platforms
 - Two heteroplasmies in two components, not in previous certificates
- Planned FSIG short communication
- Certificate update Late 2014




Site 1,393 (G/A)



AmpliSeq Identity Panel


Ion Torrent PGM







- 90 autosomal SNPs
 - Pakstis 2010
 - Kidd 2012
 - SNPforID
- 30 Y-SNPs
 - Upper clade branches









Degraded DNA / Sensitivity Study

Assays	PGM IISNPs	MiniFiler	IdentiFiler Plus	GlobalFiler	DIPlex
DEGRADED DNA SAMPLES					
50-250	X	X		X	X
50-200	X	X		X	X
50-150	X	X		X	X
50-100	X	X		X	X
50-75	X	X		X	X
SENSITIVITY SAMPLES					
0.5 ng	X		X	X	X
0.1 ng	X		X	X	X
0.05 ng	X		X	X	X









Parsing STRait Razor Output with Java

Sample name	A	B	C	D	48	Locus	height ratio
A10-H-GT37900_573_L001_R1_001.fastq.STRaitRazor		CSF1PO	CSF1PO	D10S1	49	CSF1PO	0.683
A11-H-TT51422_581_L001_R1_001.fastq.STRaitRazor		11	12	50	50	D10S1248	0.954
A12-H-Z780786_589_L001_R1_001.fastq.STRaitRazor		10	12	51	51	D12S391	N/A
A2-C-UT57318_59_L001_R1_001.fastq.STRaitRazor		10	11	52	52	D13S317	0.985
A3-C-WA29594_525_L001_R1_001.fastq.STRaitRazor		11	13	53	53	D16S539	0.961
A4-AA-JT51471_533_L001_R1_001.fastq.STRaitRazor		11	12	54	54	D18S51	0.799
A6-AA-OT05897_541_L001_R1_001.fastq.STRaitRazor		11	12	55	55	D19S433	0.983
A7-AA-PT84223_549_L001_R1_001.fastq.STRaitRazor		12	13	56	56	D151656	0.845
A8-AA-PT84232_557_L001_R1_001.fastq.STRaitRazor		12	12	57	57	D21S11	0.941
A9-H-GT37778_565_L001_R1_001.fastq.STRaitRazor		9	10	58	58	D22S1045	0.847
B1-2391c-E_S2_L001_R1_001.fastq.STRaitRazor		10	12	59	59	D251338	0.752
B10-H-GT37913_574_L001_R1_001.fastq.STRaitRazor		9	10	60	60	D25441	N/A
B11-H-TT51435_582_L001_R1_001.fastq.STRaitRazor		10	11	61	61	D3S1358	N/A
B12-H-Z780815_590_L001_R1_001.fastq.STRaitRazor		10	12	62	62	D5S818	N/A
B2-C-WT51342_S10_L001_R1_001.fastq.STRaitRazor		11	12	63	63	D7S820	0.92
B3-C-WT51373_S18_L001_R1_001.fastq.STRaitRazor		9	12	64	64	D8S1179	0.925
B4-C-WA29612_S26_L001_R1_001.fastq.STRaitRazor		10	13	65	65	FGA	0.989
B5-AA-JT51499_534_L001_R1_001.fastq.STRaitRazor		10	12	66	66	PentaD	0.935
		11	12	67	67	PentaE	N/A
		11	12	68	68	TH01	N/A
		11	12	69	69	TPOX	N/A
		70		70	70	VWA	0.787



SRM 2391c: PCR-Based DNA Profiling Standard

- Components A through D: DNA extracts in liquid form
- Components E and F: DNA spotted on 903 / FTA paper
- Certified values for STR alleles based on CE length polymorphisms

SRM 2391c
Current price: \$626 USD

Lab 1 ↔ Lab 2

Helps meet QAS Std. 9.5.5 and ISO 17025

Standard Reference Material

Calibration with SRMs enables confidence in comparisons of results between laboratories

Genomic DNAs characterized for the expanded CODIS core loci and Y-STRs

Current Values for STR Loci

Certified Values			Reference Values			Information Values		
Autosomal STR (24)	Y STR (17)	X STR (0)	Autosomal STR (23)	Y STR (0)	X STR (0)	Autosomal STR (1)	Y STR (0)	X STR (0)
D151656	DYS19	None	D1GAT A113	None	None	Penta C	None	None
D251338	DYS385a		D1S1627					
D2S441	DYS385b		D1S1677					
D3S1358	DYS389I		D2S1776					
D5S818	DYS389II		D3S3053					
D7S820	DYS390		D3S4529					
D8S1179	DYS391		D4S2364					
D8S1115	DYS392		D4S2408					
D10S1248	DYS393		D5S2500					
D12S391	DYS437		D6S1017					
D13S317	DYS438		D6S474					
D16S539	DYS439		D9S1122					
D18S51	DYS448		D9S2157					
D19S433	DYS456		D10S1435					
D21S11	DYS458		D17S1301					
D22S1045	DYS635		D17S974					
CSF1PO	Y GATA H4		D18S853					
FGA			D20S1082					
Penta D			D20S462					
Penta E			F13A01					
SE33			F13B					
TH01			FESFPS					
TPOX			LPL					
VWA								

STR Typing kits and Primer Mixes

100% Concordance with all kits

↓

Kit Provider			
Thermo Fisher (12)	Promega Corp. (6)	Qiagen Inc. (2)	Primer Mieux (2)
Identifier	PowerPlex 16	ESplex	26plex
Identifier Plus	PowerPlex 16 HS	Idplex	minSTRs
NGM	PowerPlex ESX 17		
NGM Select	PowerPlex ES1 17		
COfiler	PowerPlex ES		
Profiler	PowerPlex S5		
Profiler Plus	PowerPlex Y		
Profiler Plus ID	FFFL		
SGM Plus			
SEfiler			
MiniFiler			
Yfiler			

Kit Provider			
Thermo Fisher (14)	Promega Corp. (16)	Qiagen Inc. (9)	Primer Mieux (3)
Identifier	PowerPlex 16	ESplex	26plex
Identifier Plus	PowerPlex 16 HS	Idplex	minSTRs
NGM	PowerPlex ESX 17	ESplex SE	RM Y STRs
NGM Select	PowerPlex ES1 17	ESplex SE Plus	
COfiler	PowerPlex ES	ESplex SE GOI	
Profiler	PowerPlex S5	Idplex GOI	
Profiler Plus	PowerPlex Y	Argus X-12	
Profiler Plus ID	FFL	Idplex	
SGM Plus	PowerPlex ES1 17 Pro		
SEfiler	PowerPlex ESX 17 Fast		
MiniFiler	PowerPlex ES1 17 Fast		
Yfiler	PowerPlex 19D		
GlobalFiler	PowerPlex 21		
Yfiler Plus	PowerPlex GS7		
	PowerPlex Fusion		
	PowerPlex Y22		

Updated Typing

All certified loci have been fully sequenced


Certified Values			Reference Values			Information Values		
Autosomal STR (25)	Y STR (29)	X STR (0)	Autosomal STR (23)	Y STR (0)	X STR (0)	Autosomal STR (1)	Y STR (0)	X STR (12)
D1S1636	DYS19	None	D1GATA113	None	None	Penta C	None	DXS7132
D2S1338	DYS385a		D1S1627					DXS7423
D2S441	DYS385b		D1S1677					DXS8378
D3S1358	DYS389I		D2S1776					DXS10074
D5S818	DYS389II		D3S3053					DXS10079
D6S1043	DYS390		D3S4529					DXS10101
D7S820	DYS391		D4S2364					DXS10103
D8S1179	DYS392		D4S2408					DXS10134
D8S1115	DYS393		D5S2500					DXS10135
D10S1248	DYS437		D6S1017					DXS10146
D12S391	DYS438		D6S474					DXS10148
D13S317	DYS439		D9S1122					HPRTB
D16S539	DYS448		D9S2157					
D18S51	DYS456		D10S1438					
D19S433	DYS458		D17S1301					
D21S11	DYS635		D17S974					
D22S1045	Y GATA H4		D18S853					
CSP1PO	DYS449		D20S1082					
FGA	DYS460		D20S482					
Penta D	DYS481		F13A01					
Penta E	DYS518		F13B					
SE33	DYS533		FESFPS					
TH01	DYS549		LPL					
TPOX	DYS570							
vWA	DYS576							
	DYS627							
	DYS643							
	DYF387S1a							
	DYF387S1b							

Argus X-12 kit

*Other Information Value Considerations:
 - DIPplex kit loci
 - PGM and MiSeq IISNPs

New Y-STR loci in commercial kits (Yfiler Plus & PPY23)
Update to be completed by Oct. 2014

Acknowledgements



NIST
 Katherine Gettings
 Erica Butts
 Margaret Kline
 Dave Duewer
 Kevin Kiesler
 Harish Swaminathan
 Jo Lynne Harenza
 Nate Olson

Battelle
 Seth Faith
 Rich Guerrieri
 Liz Montano
 Brian Young
 Esley Heizer
 Francisco Martinez

Promega
 Doug Storts
 Jay Patel

ThermoFisher
 Nnamdi Ihuegbu
 Rob Legace
 Narasi Rajagopalan
 Wenchi Liao
 Gloria Lam
 Sharon Wootton
 Joseph Chang

Contact Information
 peter.vallone@nist.gov