

YOUR INTERNATIONAL FORENSICS HUB
ATLANTA, GA • OCT. 7-10, 2013



ISHI

INTERNATIONAL SYMPOSIUM
ON HUMAN IDENTIFICATION

ISHI Workshop on New Loci and Kits

October 10, 2013 (Atlanta, GA)

New Autosomal and Y-STR Loci and Kits:

Making Data Driven Decisions

Experience with PowerPlex Fusion

Jeffrey Nye

Michigan State Police



Michigan State Police

STR Kit Chemistry History

- Transitioned from RFLP directly to STRs
- Initial casework kits included AMPFLSTR Blue, Green I and II
 - Provided 9 STR Loci
- Casework moved to Profiler Plus/Cofiler
 - Provided 13 STR Loci
- Database group utilized PowerPlex 16
 - Provided 13 STR Loci plus Penta E and Penta D
- Casework moved to PowerPlex 16 HS in 2011
 - Interest in capturing Penta E and Penta D markers
 - Detailed comparison of Identifiler Plus and PowerPlex 16 HS

PowerPlex Fusion



- Interest in looking at extended STR Loci panels
- Currently use 3130 and 3130xl Genetic Analyzer platforms
 - Contracts to purchase 3500s being established
- Offender database of 300,000+ include Penta loci

• INTERNAL VALIDATION

- Casework and Database units
- Studies completed include:
 - Precision
 - Sensitivity
 - Baseline noise evaluation
 - Contamination Assessment
 - Mock Casework
 - Mixtures
 - Concordance
 - NIST SRM

Validation at One Laboratory
(Recently Completed)

Performance Verification
Studies at Two Casework
Laboratories and Our
Databasing Laboratory

Training, Competency Testing
and Planned Implementation
System-wide.

Internal Validation



Precision

Ladders were run multiple times over a few days. All capillaries were tested.

A table in GMID was created and exported as a tab-delimited file for import into Microsoft Excel

The average bp size and standard deviation were calculated for each allele of each locus



Internal Validation



Precision

In general:

Smaller loci are averaging differences approximately 0.03 to 0.05bp

Larger loci (FGA, PentaE, PentaD) are averaging 0.08-0.09bp

Well below 0.5bp window





Internal Validation



Contamination Assessment and Baseline Noise Evaluation

40 Reagent Blanks were amplified

- Stain Extraction Blanks
 - Maxwell Extraction Blanks
 - Epithelial Extraction Blanks
 - Sperm Fraction Blanks
-
- Run on 3130
 - Analyzed at 5 rfu
 - GMID table created and exported to Excel
 - Average Peak height was 5.8rfu +/- 3.2 rfu
 - A few peaks were labeled > 25 rfu
- 



Internal Validation

Sensitivity

3 samples of extracted DNA were used
(Many heterozygous loci)

Samples were quantified using Plexor HY

Serial dilutions were made

Each dilution was quantified in triplicate

Replicates were amplified. (Range of >1ng to ~ 8 pg)



Internal Validation


Sensitivity

Samples less than 0.5 ng were amplified
12 times each



Data was reviewed for imbalance and dropout

Full profiles consistently down to 125pg


Increase in concentration = better PHR




Internal Validation



Internal Validation





0.125ng



0.06ng

Internal Validation



Internal Validation



Mixtures



Mixtures prepared from extracted reference sample DNA

Ratios from 100:0 to 0:100 were prepared with an input template amount of 0.75ng

Amplified product was ran on a 3130 with a standard 5 second injection time

Minor donor detected at all mixture ratios





Internal Validation

Casework

Fortunate to be able to use actual casework samples

More than 40 cases have been analyzed with Fusion-
only able to show a small portion here


Concordant data overall






Description	Quantification Value
Left Hand Fingernail Scrapings	0.0304
Right Hand Fingernail Scrapings	0.1650
swabbings from gray hat	0.0074
swab of piece of glove	0.0702
swab of water bottle	0.3330
swab of Miller Lite bottle	0.0118
swab of hat	1.0500
swab of Gatorade bottle	0.0077
swab of Stanley tool	0.0071
swab of glove (under TV)	0.5710
swab of glove (parking lot)	1.8300
Cigarette Butt	0.1210
Cigarette Butt	0.0916
Leather Apron - Lower Left	0.3520
Leather Apron - Lower Left Corner	0.1330
Leather Apron - Upper Center	0.0064
swab from inside stairs	1.1900
cigarette	2.4200
blood from jewelry box	0.5510
blood from kitchen table	0.2680
blood from glass	0.1160
cigarette butt (8A)	0.5890
cigarette butt (CPeak-1)	4.4800
blood from air bag	0.1400
blood on gift card sleeve	0.1150

Description	Quantification Value
swab of cords	0.1730
swab of bottle	0.5870
swab of sweatshirt	0.5490
swab of mouth of water bottle	3.2800
blood from camp fuel can	0.7290
envelope cutting	0.3710
blood from glass	0.5560
suspected flesh	5.1500
cigarette butt	2.9100
blood from POE	0.4110
blood from lottery slip	0.1550
blood from broken window	0.6800
blood from computer bag	0.1840
blood from broken window	0.3430
blood from floor	4.0500


Description	Quantification Value	Y Quant
Vaginal / Cervical - EF	1.1200	0.0075
Anal / Rectal - EF	1.3400	0.0250
Vaginal / Cervical - EF	1.1900	N/A
Vaginal - EF	0.6990	0.0940
Vaginal - EF	1.0900	N/A
Vaginal / Cervical - SF	2.3900	5.0700
Anal / Rectal - SF	0.1570	0.3490
Vaginal / Cervical - SF	10.3000	0.0080
Vaginal - SF	9.3700	9.3800
Vaginal - SF	6.8100	0.0346








Swab of a water bottle



D5S818 D13S317 D7S820 D16S539 CSF1PO Penta_D

80 160 240 320 400 480

PP16HS




Swab of Beer Bottle

D16S539 D18S51 D2S1338 CSF1PO Penta D

70 140 210 280 350 420 490

Fusion




D3S1358


TH01

D21S11


D18S51

Penta_E







Fusion: Swab of Hat






PP16HS: Fingernail Scrapings

5 total markers interpretable



Fusion: Fingernail Scrapings



10 total markers interpretable



• INTERNAL VALIDATION

- NIST Standards Concordance

- NIST sample set 2931b was amplified with Fusion chemistry.
- All profiles obtained were concordant with the published profiles for the various components provided in the kit.




• INTERNAL VALIDATION

• NIST Standards Concordance


• 2800M

Sample Name	Marker	Allele 1	Allele 2	Certified Value
2800M071713 [3 sec]	CSF1PO	12		12,12
2800M071713 [3 sec]	D3S1358	17	18	17,18
2800M071713 [3 sec]	D5S818	12		12,12
2800M071713 [3 sec]	D7S820	8	11	8,11
2800M071713 [3 sec]	D8S1179	14	15	14,15
2800M071713 [3 sec]	D13S317	9	11	9,11
2800M071713 [3 sec]	D16S539	9	13	9,13
2800M071713 [3 sec]	D18S51	16	18	16,18
2800M071713 [3 sec]	D21S11	29	31.2	29,31.2
2800M071713 [3 sec]	FGA	20	23	20,23
2800M071713 [3 sec]	TH01	6	9.3	6,9.3
2800M071713 [3 sec]	TPOX	11		11,11
2800M071713 [3 sec]	vWA	16	19	16,19
2800M071713 [3 sec]	AMEL	X	Y	X,Y
2800M071713 [3 sec]	Penta D	12	13	12,13
2800M071713 [3 sec]	Penta E	7	14	7,14
2800M071713 [3 sec]	D2S1338	22	25	22,25
2800M071713 [3 sec]	D19S433	13	14	13,14
2800M071713 [3 sec]	D10S1248	13	15	13,15
2800M071713 [3 sec]	D2S441	10	14	10,14
2800M071713 [3 sec]	D22S1045	16		16,16
2800M071713 [3 sec]	D12S391	18	23	18,23
2800M071713 [3 sec]	D1S1656	12	13	12,13
2800M071713 [3 sec]	DYS391	10		10,10



• INTERNAL VALIDATION


- NIST Standards Concordance
 - Genomic DNA 3



Sample Name	Marker	Allele 1	Allele 2	Allele 3	Certified Value
303 - Genomic DNA 3 [3 sec]	CSF1PO	11	12		11,12
303 - Genomic DNA 3 [3 sec]	D3S1358	9 (95 RFU = 2.7%)	15		15,15
303 - Genomic DNA 3 [3 sec]	D5S818	11			11,11
303 - Genomic DNA 3 [3 sec]	D7S820	12	13		12,13
303 - Genomic DNA 3 [3 sec]	D8S1179	14	15 (287 RFU = 12.1%)	16	14,16
303 - Genomic DNA 3 [3 sec]	D13S317	11	12		11,12
303 - Genomic DNA 3 [3 sec]	D16S539	11	12		11,12
303 - Genomic DNA 3 [3 sec]	D18S51	16	20		16,20
303 - Genomic DNA 3 [3 sec]	D21S11	28	31.2		28,31.2
303 - Genomic DNA 3 [3 sec]	FGA	23	25		23,25
303 - Genomic DNA 3 [3 sec]	TH01	9.3			9.3,9.3
303 - Genomic DNA 3 [3 sec]	TPOX	8	11		8,11
303 - Genomic DNA 3 [3 sec]	vWA	18	19		18,19
303 - Genomic DNA 3 [3 sec]	AMEL	X	Y		X,Y
303 - Genomic DNA 3 [3 sec]	Penta D	11	12		11,12
303 - Genomic DNA 3 [3 sec]	Penta E	13	14		13,14
303 - Genomic DNA 3 [3 sec]	D2S1338	20	24		20,24
303 - Genomic DNA 3 [3 sec]	D19S433	12	14		12,14
303 - Genomic DNA 3 [3 sec]	D10S1248	13	16		13,16
303 - Genomic DNA 3 [3 sec]	D2S441	10	14		10,14
303 - Genomic DNA 3 [3 sec]	D22S1045	15	16		15,16
303 - Genomic DNA 3 [3 sec]	D12S391	15	21		N/A
303 - Genomic DNA 3 [3 sec]	D1S1656	14	15		N/A
303 - Genomic DNA 3 [3 sec]	DYS391	10			N/A

• INTERNAL VALIDATION

- NIST Standards Concordance
- Genomic DNA 10 (GM09948)




Sample Name	Marker	Allele 1	Allele 2	Allele 3	Certified Value
310 - Genomic DNA 10 [3 sec]	CSF1PO	10	11	12	10,11,12
310 - Genomic DNA 10 [3 sec]	D3S1358	15	17		15,17
310 - Genomic DNA 10 [3 sec]	D5S818	11	13		11,13
310 - Genomic DNA 10 [3 sec]	D7S820	11			11,11
310 - Genomic DNA 10 [3 sec]	D8S1179	12	13		12,13
310 - Genomic DNA 10 [3 sec]	D13S317	11			11,11
310 - Genomic DNA 10 [3 sec]	D16S539	11			11,11
310 - Genomic DNA 10 [3 sec]	D18S51	15	18		15,18
310 - Genomic DNA 10 [3 sec]	D21S11	29	30		29,30
310 - Genomic DNA 10 [3 sec]	FGA	24	26		24,26
310 - Genomic DNA 10 [3 sec]	TH01	6	9.3		6,9.3
310 - Genomic DNA 10 [3 sec]	TPOX	8	9		8,9
310 - Genomic DNA 10 [3 sec]	vWA	17			17,17
310 - Genomic DNA 10 [3 sec]	AMEL	X	Y		X,Y
310 - Genomic DNA 10 [3 sec]	Penta D	8	12		8,12
310 - Genomic DNA 10 [3 sec]	Penta E	11			11,11
310 - Genomic DNA 10 [3 sec]	D2S1338	23			23,23
310 - Genomic DNA 10 [3 sec]	D19S433	13	14		13,14
310 - Genomic DNA 10 [3 sec]	D10S1248	12	15		12,15
310 - Genomic DNA 10 [3 sec]	D2S441	11	12		11,12
310 - Genomic DNA 10 [3 sec]	D22S1045	16	17 (225 RFU = 16.9%)	18	16,18
310 - Genomic DNA 10 [3 sec]	D12S391	18	24	24.3 (112 RFU = 4.5%)	N/A
310 - Genomic DNA 10 [3 sec]	D1S1656	14	17		N/A
310 - Genomic DNA 10 [3 sec]	DYS391	10			N/A

• INTERNAL VALIDATION



Direct Amplification of Knowns

- Vary amount of swab
 - Cotton Swab – 1 swab, ½ swab, ¼ swab
 - Omni Swab – 2 teeth, 1 tooth, ½ tooth
- Cycle Number Optimization
 - 30 Cycles / 27 Cycles / 25 Cycles
- Procedure:
 - Add 1 mL of swab solution to swab amounts listed above.
 - Place samples in 70 C hot plate for 30 minutes.
 - Use 1 uL of solution for amp 1 and 2 uL solution for amp 2.
 - Amp samples at same cycle / extension as evidence cycle parameters. (30 cycles)
 - Additional Amps performed at 27, 26, and 25 Cycles



INTERNAL VALIDATION

Direct Amplification of Knowns







Artifact Assessments



• Additional Studies

- Comparison of artifacts



Additional Studies

Comparison of artifacts

D18 WITH PP16HS

D18S51

300

12
299.38
595

17
319.13
536

18
323.01
483

artifacts -
up of loci

D18 WITH FUSION

D18S51


165

17
607
171.63

18
550
175.66

• Additional Studies



- Comparison of artifacts



Anal Swab
PP16HS

• Additional Studies

- Comparison of artifacts



Considerations for Selecting a Kit



- One kit vs. two kits
- 5 dye vs. 6 dye
- Familiarity

- Software changes
- Ease of Direct Amplification
- Support

- Use of Pentas
- Cost
- Required CODIS loci (No changes yet)

ACKNOWLEDGEMENTS

Thank you to Promega Corporation for asking us to participate and all of the other collaborators on the project.



A special thanks to Kirk DeLeeuw, Josh Strong, Donald Yet and Kristin Schelling for all of the effort in conducting the internal validation.